

# **Extra INSTAND EQA Scheme (340) - April 2020**

## **Virus Genome Detection SARS-CoV-2**

**Interim Evaluation of Results**

**Submitted by 112 out of 484 Laboratories  
(as of 14 April 2020, 01:57 pm Berlin time)**

**Heinz Zeichhardt<sup>1,2,3,4</sup>, Martin Kammel<sup>2,3</sup> and Hans-Peter Grunert<sup>4</sup>**

<sup>1</sup>Professor für Virologie (i.R.)  
Charité - Universitätsmedizin Berlin

<sup>2</sup>INSTAND e.V. - Gesellschaft zur Förderung der Qualitätssicherung  
in medizinischen Laboratorien e.V, Düsseldorf

<sup>3</sup>IQVD GmbH - Institut für Qualitätssicherung in der Virusdiagnostik

<sup>4</sup>GBD Gesellschaft für Biotechnologische Diagnostik mbH, Berlin

# **Disclosure**

Heinz Zeichhardt is share-holder  
of

GBD Gesellschaft für Biotechnologische Diagnostik mbH, Berlin  
and

IQVD GmbH - Institut für Qualitätssicherung in der Virusdiagnostik, Berlin

# EQA Network

## Scientific umbrella

German Association for Prevention of Virus Diseases (DVV e.V.)

Society of Virology (GfV e.V.)

German Society for Hygiene and Microbiology (DGHM e.V.)

## Partners

### Berlin

Teams of

GBD Gesellschaft für Biotechnologische  
Diagnostik mbH, Berlin

and

IQVD GmbH - Institut für Qualitätssicherung  
in der Virusdiagnostik, Berlin

and

38 INSTAND Expert Laboratories  
incl.  
Robert Koch-Institut  
Paul-Ehrlich-Institut  
National Reference and Consiliary Labs

and

**EMPIR**  
**AntiMicroResist**

**EMRP**  
European Metrology Research Programme  
►Programmes of EURAMET  
The EMRP is jointly funded by the EMRP participating countries  
within EURAMET and the European Union



**EURAMET**  
European Association of National Metrology Institutes

and

### Düsseldorf

INSTAND-Team

JRC – Joint Research Centre (BE)




LGC – National Measurement Laboratory (UK)

NIB – National Institute of Biology (SI)


PTB - Physikalisch-Technische Bundesanstalt (DE)

# The Cycle of Laboratory Diagnosis



Quality Management  
in Medical Laboratories  
ISO 9000:2000  
Wemmie Elsenga  
[http://www.epbs.net/brussels/b  
russels03\\_elsenga.htm](http://www.epbs.net/brussels/bRussels03_elsenga.htm)

# The Role of Quality Control for Public Health and Quality Improvement of Diagnostics and Blood Safety



# INSTAND e.V. - Quality Assurance in (Virus) Diagnostics

mandated and harmonized  
standard  
ISO EN DIN 15189

in vitro Diagnostic  
Medical Device Directive 98/79  
EC

European law  
**REGULATION (EU) 2017/746  
OF THE EUROPEAN PARLIAMENT  
AND OF THE COUNCIL** of 5 April 2017  
on in vitro diagnostic medical devices  
and repealing Directive 98/79/EC and  
Commission Decision 2010/227/EU

Medical Devices Act  
(Medizinproduktegesetz/MPG)

German law

RiliBÄK  
Bundesärztekammer (BÄK)

Guideline of  
German Medical Association

Scientific societies  
DVV\* and GfV\*\*  
**Joint Diagnostic Council**

EQA schemes  
**INSTAND e.V.**  
scientific society  
cooperating with scientific societies

EQA  
schemes  
RfB

WHO

Federal institutes  
(RKI/CDC + PEI/FDA)

**ICBS**  
International Consortium for Blood Safety

38 International and  
National Reference Centers  
and Consiliary Laboratories  
and  
Target Value Laboratories

**IQVD GmbH**  
+  
**GBD mbH**

\* DVV; German Association for the Control of Virus Diseases  
("Deutsche Vereinigung zur Bekämpfung der Viruskrankheiten e.V.")

\*\* GfV; Society of Virology  
("Gesellschaft für Virologie e.V.")



## INSTAND E.V. – A SCIENTIFIC SOCIETY COOPERATING WITH INTERNATIONAL AND NATIONAL SOCIETIES AND ORGANISATIONS

### International

14 Organisations / Scientific Societies

#### INTERNATIONAL

- Clinical and Laboratory Standards Institute (CLSI, formerly NCCLS)
- EQAS/PT-Provider, QualiCont, Hungary
- European Committee for Standardization (CEN)
- European Organisation For External Quality Assurance Providers in Laboratory Medicine (EQALM)
- International Consortium for Blood Safety (ICBS)
- International Organisation for Standardisation (ISO)
- Joint Committee for Traceability in Laboratory Medicine (JCTLM)
- MQ Zürich Schweiz
- National Institute of Standards and Technology (NIST)
- Österreichische Gesellschaft für Qualitätssicherung und Standardisierung medizinisch-diagnostischer Untersuchungen (ÖQUASTA)
- Working Group HbA1c (IFCC)
- Working Group HbA2 (IFCC)
- World Association of Pathology and Laboratory Medicine (WASPoLM)
- World Health Organisation

### National

28 Organisations / Scientific Societies


#### NATIONAL

- Arbeitsgemeinschaft der Wissenschaftlichen Medizinischen Fachgesellschaften e.V. (AWMF)
- Arbeitsgruppe Therapeutisches Drug Monitoring (AGTDM) der Arbeitsgemeinschaft Neuropsychopharmacologie und Pharmakopsychiatrie e.V. (AGNP e.V.)
- Arbeitskreis Mykobakterien (AKM)
- Ärzteverband Deutscher Allergologen e.V. (AeDA)
- Berufsverband Deutscher Dermatologen e.V. (BVDD)
- Bundesärztekammer (BÄK, German Medical Association)

- Deutsche Akkreditierungsstelle (DAkkS)
- Deutsche Dermatologische Gesellschaft e.V. (DDG)
- Deutsche Gesellschaft für Hämatologie und Onkologie e.V. (DGHO)
- Deutsche Gesellschaft für Hygiene und Mikrobiologie e.V. (DGHM)
- Deutsche Gesellschaft für Immunogenetik e.V. (DGI)
- Deutsche Gesellschaft für Innere Medizin e.V. (DGIM)
- Deutsche Gesellschaft für Internationale Zusammenarbeit (GIZ)
- Deutsche Gesellschaft für Liquordiagnostik und Klinische Neurochemie e.V. (DGLN)
- Deutsche Gesellschaft für Mineralstoffe und Spurenelemente e.V. (GMS)
- Deutsche Gesellschaft für Parasitologie e.V. (DGP)
- Deutsche Gesellschaft für Transfusionsmedizin und Immunhämatologie e.V. (DGTI)
- Deutsche Gesellschaft für Urologie e.V. (GU)
- Deutsche Gesellschaft für Zytologie e.V. (DGZ)
- Deutsche Vereinigung zur Bekämpfung der Viruskrankheiten e.V. (DVV)
- Deutsche Vereinte Gesellschaft für Klinische Chemie und Laboratoriumsmedizin e.V. (DGKL)
- Deutschsprachige Mykologische Gesellschaft e.V. (DMykG)
- DIN - Deutsches Institut für Normung with special activities in „Normenausschuss Medizin“
- Gesellschaft für Thrombose und Hämostaseforschung e.V. (GTH)
- Gesellschaft für Virologie e.V. (GIV)
- National organisations for technical assistants and continuous education (dvta and diw-mta e.V.)
- Physikalisch-Technische Bundesanstalt (PTB)
- Zentralstelle der Länder für Gesundheitsschutz bei Arzneimitteln und Medizinprodukten (ZLG)

# INSTAND e.V.

## Cooperation with participants from more than 80 Countries



# INSTAND e.V.

Immunohaematology  
(41)

Autoimmune diseases (57)

Haemostasis  
(54)

Haematology  
(54)

Clinical chemistry  
(43)

Trace elements  
(50)

Pharmaceutics and drugs  
(152)

> 350  
Programs

Virtual EQA schemes  
(3)

Virology  
(78)

Bacteriology / serology  
(47)

Bacterial genome detection  
(17)

Parasitology  
(8)

Mycology  
(7)

Molecular diagnostics  
(90)

POCT  
(34)

EQAS in virus diagnostics - 2019:  

- > 1700 participating laboratories
- from > 60 countries worldwide

# INSTAND EQA Schemes (78 Schemes) in Virus Immunology and Genome Detection 2020

Serology and Antigen Detection		Virus Genome Detection and Typing		
HIV-1/2	Herpes simplex Viruses	HIV-1 (RNA) HIV-2 (RNA)	Measles Rubella Mumps Viruses	
HIV-1 p24 Ag	Varicella Zoster Virus	HIV-1 Resistance + Tropism		Adenoviruses
HTLV-1/2	Epstein Barr Virus	Hepatitis A Virus		Norovirus Rotavirus
Hepatitis A Virus	Resp. Sync. Virus Ag	Hepatitis B Virus + Genotyping + Resistance	CoV incl. MERS CoV <b>SARS-CoV-2 NEW</b>	
Hepatitis B Virus, Prg. I	Influenza A and B Ag + A/H1N1 pdm 2009 + A/H5N1+A/H7N9	Hepatitis C Virus + Genotyping + Resistance	Enteroviruses + Enterovirus (WHO/RKI) Parechovirus <b>NEW</b>	
Hepatitis B Virus, Prg. II	Rubella Virus Measles Virus Mumps Virus	Hepatitis D Virus Hepatitis E Virus	Human Rhinoviruses Resp. Syncytial Virus Hum. Metapneumovirus Parainfluenza viruses	
Hepatitis C Virus	TBE Virus	Torque Teno Virus <b>NEW</b>	Influenza A and B incl avian	
Hepatitis D Virus	Hantavirus	Cytomegalovirus + Resistance	BK Virus	JC Virus
Hepatitis E Virus	Dengue Virus Chikungunya Virus Zika Virus	Epstein-Barr Virus Herpes simplex Viruses HHV-6 <b>NEW</b> HHV-8 <b>NEW</b>	Dengue Virus West Nile Virus Chikungunya Virus Zika Virus	
Parvovirus B19	Rabies Virus	Varicella Zoster Virus	Hum. Papilloma Viruses	
Cytomegalovirus	Borna Virus <b>NEW</b>	Parvovirus B19	Rabies Virus	
<b>SARS-CoV-2 NEW</b>	BSE (PrPsc) (2002-2007)	Multiplex: Respiratory Viruses (2 programs)	Multiplex: <b>NEW</b> Viral Meningitis/Encephalitis	
		Multiplex: Gastrointestinal Viruses	Borna Viruses <b>NEW</b>	

# **Extra INSTAND EQA Scheme (340) - April 2020**

## **Virus Genome Detection SARS-CoV-2**

### **Cooperation partner**

National Consultant Laboratory for Coronaviruses

Institute of Virology, Charité – University Medicine Berlin, Campus Charité Mitte

Prof. Dr. Christian Drosten

Dr. Victor M. Corman

Dr. Daniela Niemeyer

### **Expert Laboratories**

Universitätsklinikum Frankfurt,  
Institut für Medizinische Virologie  
Prof. Dr. Sandra Ciesek  
Prof. Dr. Holger F. Rabenau  
Prof. Dr. Annemarie Berger

Uniklinik Köln, Institut für Virologie  
Nationales Referenzzentrum für  
Papillom- und Polyomaviren  
Prof. Dr. Florian Klein  
Prof. Dr. Ulrike Wieland  
Dr. Steffi Silling  
Dr. Rolf Kaiser  
Dr. Eva Heger  
Dr. Elena Knops

Medizinisches Infektionszentrum  
Berlin  
Dr. M. Obermeier  
Dr. R. Ehret

# Interim Evaluation

## Based on the Results of 112 out of 484 Laboratories

Reported by 14 April 2020 01:57 pm Berlin Time

Sample no.	Sample property
340059	SARS-CoV-2 / 1 : 1 000 diluted\$
340060	CoV OC43 / 1 : 2 500 diluted / specificity control
340061	???
340062	???
340063	???
340064	SARS-CoV-2 / 1 : 100 000 diluted\$
340065	???

\$ The SARS-CoV-2 positive samples 340059 and 340064 represent different dilution steps from a dilution series of a lysate from cells, infected with SARS-CoV-2 (inactivated).

# Interim Evaluation

## Based on the Results of 112 out of 484 Laboratories

Reported by 14 April 2020 01:57 pm Berlin Time

Sample No.	Sample properties	Expected qualitative result for SARS-CoV-2	Gene region	Correct results per reported results differentiated by gene region	Reported Ct/Cp-results differentiated by gene region median (min – max)
Sample 340059	SARS-CoV-2 1 : 1 000 diluted <sup>\$</sup>	positive	E	83/83 (100%)	22.3 (17.7-29.4)
			N	22/22 (100%)	22.9 (17.9-30.9)
			ORF1a	7/7 (100%)	22.8 (21.7-28.7)
			ORF1ab	1/1 (100%)	20.9 (20.9-20.9)
			RdRP	41/41 (100%)	23.5 (20.5-34.4)
			S	27/27 (100%)	21.7 (18.0-24.8)
			n.s. <sup>§</sup>	29/29 (100%)	22.2 (10.4-33.0)
			total	210/210 (100%)	22.6
Sample 340064	SARS-CoV-2 1 : 100 000 diluted <sup>\$</sup>	positive	E	83/83 (100%)	29.1 (24.1-33.5)
			N	22/22 (100%)	29.6 (26.2-37.5)
			ORF1a	7/7 (100%)	28.7 (28.0-30.3)
			ORF1ab	1/1 (100%)	29.6 (29.6-29.6)
			RdRP	41/41 (100%)	30.4 (26.9-41.4)
			S	27/27 (100%)	28.6 (25.0-31.5)
			n.s. <sup>§</sup>	29/29 (100%)	29.3 (16.0-40.0)
			total	210/210 (100%)	29.4
Sample 340060	CoV OC43 1 : 2 500 diluted specificity control*	negative	E	78/83 (94.0%)	-
			N	19/21 (90.5%) <sup>#</sup>	-
			ORF1a	7/7 (100%)	-
			ORF1ab	1/1 (100%)	-
			RdRP	36/39 (92.3%) <sup>#</sup>	-
			S	27/27 (100%)	-
			n.s. <sup>§</sup>	29/29 (100%)	-
			total	197/210 (93.8%) <sup>#</sup>	-

# Interim Evaluation

## Based on the Results of 112 out of 484 Laboratories

Reported by 14 April 2020 01:57 pm Berlin Time

Sample No.	Sample properties	Expected qualitative result for SARS-CoV-2	Gene region	Correct results per reported results differentiated by gene region	Reported Ct/Cp-results differentiated by gene region median (min – max)
Sample 340059	<b>SARS-CoV-2 1 : 1 000 diluted<sup>\$</sup></b>	<b>positive</b>	E	83/83 (100%)	22.3 (17.7-29.4)
			N	22/22 (100%)	22.9 (17.9-30.9)
			ORF1a	7/7 (100%)	22.8 (21.7-28.7)
			ORF1ab	1/1 (100%)	20.9 (20.9-20.9)
			RdRP	41/41 (100%)	23.5 (20.5-34.4)
			S	27/27 (100%)	21.7 (18.0-24.8)
			n.s. <sup>§</sup>	29/29 (100%)	22.2 (10.4-33.0)
		<b>total</b>		<b>210/210 (100%)</b>	<b>22.6</b>
Sample 340064	<b>SARS-CoV-2 1 : 100 000 diluted<sup>\$</sup></b>	<b>positive</b>	E	83/83 (100%)	29.1 (24.1-33.5)
			N	22/22 (100%)	29.6 (26.2-37.5)
			ORF1a	7/7 (100%)	28.7 (28.0-30.3)
			ORF1ab	1/1 (100%)	29.6 (29.6-29.6)
			RdRP	41/41 (100%)	30.4 (26.9-41.4)
			S	27/27 (100%)	28.6 (25.0-31.5)
			n.s. <sup>§</sup>	29/29 (100%)	29.3 (16.0-40.0)
		<b>total</b>		<b>210/210 (100%)</b>	<b>29.4</b>
Sample 340060	<b>CoV OC43 1 : 2 500 diluted specificity control*</b>	<b>negative</b>	E	78/83 (94.0%)	-
			N	19/21 (90.5%) <sup>#</sup>	-
			ORF1a	7/7 (100%)	-
			ORF1ab	1/1 (100%)	-
			RdRP	36/39 (92.3%) <sup>#</sup>	-
			S	27/27 (100%)	-
			n.s. <sup>§</sup>	29/29 (100%)	-
		<b>total</b>		<b>197/210 (93.8%)<sup>#</sup></b>	-

# Interim Evaluation

## Based on the Results of 112 out of 484 Laboratories

Reported by 14 April 2020 01:57 pm Berlin Time

Sample: 340064 – SARS-CoV-2 positive (inactivated), 1 : 100 000 diluted

Reagent	Testkit	Total (N=210)	positive (N=210)	Quota (100%)	Ct/Cp Median	Ct/Cp Min	Ct/Cp Max
<b>E-Gene</b>							
ALTONA DIAGNOSTIC	RealStar SARS-CoV-2 RT-PCR Kit 1.0	14	14	100,00%	29.1	26.5	32.5
OTHER MANUFACTURERS		2	2	100,00%			
IN HOUSE		7	7	100,00%	30.1	27.6	33.5
ELITech	GeneFinder COVID-19 Plus	2	2	100,00%	28.1	26.9	29.2
EURO IMMUN	EURORealTime SARS-CoV-2	1	1	100,00%	30.2	30.2	30.2
MIKROGEN	amplicube Coronavirus SARS-CoV-2 RUO	1	1	100,00%			
Priv. Inst. f. Immunol. u. Mol.genetik	AmpliGnost CoV-2 E-Gen	3	3	100,00%	26.2	25.5	29.5
QIAGEN	QIAstat-Dx Respiratory 2019-nCoV Panel	2	2	100,00%	30.9	30.8	30.9
R-BIOPHARM	RIDA GENE SARS-CoV-2 RUO	10	10	100,00%	29.7	24.1	32.8
ROCHE DIAGNOSTICS	COBAS SARS-CoV-2 Test	13	13	100,00%	29.0	28.5	32.0
SEEGENE	Allplex 2019 n-CoV Assay	9	9	100,00%	27.9	26.7	30.7
TIB MOLBIOL	LightMix Modular SARS and Wuhan CoV E-gene	19	19	100,00%	29.4	27.3	31.5
		83	83	100,00%	29.05	24.1	33.5
<b>N-Gene</b>							
OTHER MANUFACTURERS		3	3	100,00%	26.2	26.2	26.2
APPLIED BIOSYSTEMS	TaqMan 2019-nCoV Assay Kit v1	1	1	100,00%	26.3	26.3	26.3
IN HOUSE		2	2	100,00%	29.1	28.0	30.1
ELITech	GeneFinder COVID-19 Plus	2	2	100,00%	27.9	26.2	29.5
EURO IMMUN	EURORealTime SARS-CoV-2	1	1	100,00%	27.8	27.8	27.8
PATHOFINDER	RealAccurate Quadruplex Corona-plus PCR Kit	1	1	100,00%	31.4	31.4	31.4
SEEGENE	Allplex 2019 n-CoV Assay	9	9	100,00%	29.6	28.9	31.4
TIB MOLBIOL	LightMix Modular SARS and Wuhan CoV N-gene	3	3	100,00%	37.3	35.0	37.5
		22	22	100,00%	29.6	26.2	37.5
<b>ORF1a</b>							
MIKROGEN	amplicube Coronavirus SARS-CoV-2 RUO	1	1	100,00%			
ROCHE DIAGNOSTICS	COBAS SARS-CoV-2 Test	6	6	100,00%	28.7	28.0	30.3
		7	7	100,00%	28.7	28.0	30.3
<b>ORF1ab</b>							
APPLIED BIOSYSTEMS	TaqMan 2019-nCoV Assay Kit v1	1	1	100,00%	29.6	29.6	29.6
		1	1	100,00%	29.6	29.6	29.6

# Interim Evaluation

## Based on the Results of 112 out of 484 Laboratories

Reported by 14 April 2020 01:57 pm Berlin Time

Sample: 340064 – SARS-CoV-2 positive (inactivated), 1 : 100 000 diluted

Reagent	Testkit	Total (N=210)	positive (N=210)	Quota (100%)	Ct/Cp Median	Ct/Cp Min	Ct/Cp Max
<b>E-Gene</b>							
ALTONA DIAGNOSTIC	RealStar SARS-CoV-2 RT-PCR Kit 1.0	14	14	100,00%	29.1	26.5	32.5
OTHER MANUFACTURERS		2	2	100,00%			
IN HOUSE		7	7	100,00%	30.1	27.6	33.5
ELITech	GeneFinder COVID-19 Plus	2	2	100,00%	28.1	26.9	29.2
EURO IMMUN	EURORealTime SARS-CoV-2	1	1	100,00%	30.2	30.2	30.2
MIKROGEN	amplicube Coronavirus SARS-CoV-2 RUO	1	1	100,00%			
Priv. Inst. f. Immunol. u. Mol.genetik	AmpliGnost CoV-2 E-Gen	3	3	100,00%	26.2	25.5	29.5
QIAGEN	QIAstat-Dx Respiratory 2019-nCoV Panel	2	2	100,00%	30.9	30.8	30.9
R-BIOPHARM	RIDA GENE SARS-CoV-2 RUO	10	10	100,00%	29.7	24.1	32.8
ROCHE DIAGNOSTICS	COBAS SARS-CoV-2 Test	13	13	100,00%	29.0	28.5	32.0
SEEGENE	Allplex 2019 n-CoV Assay	9	9	100,00%	27.9	26.7	30.7
TIB MOLBIOL	LightMix Modular SARS and Wuhan CoV E-gene	19	19	100,00%	29.4	27.3	31.5
		83	83	100,00%	29.05	24.1	33.5
<b>N-Gene</b>							
OTHER MANUFACTURERS		3	3	100,00%	26.2	26.2	26.2
APPLIED BIOSYSTEMS	TaqMan 2019-nCoV Assay Kit v1	1	1	100,00%	26.3	26.3	26.3
IN HOUSE		2	2	100,00%	29.1	28.0	30.1
ELITech	GeneFinder COVID-19 Plus	2	2	100,00%	27.9	26.2	29.5
EURO IMMUN	EURORealTime SARS-CoV-2	1	1	100,00%	27.8	27.8	27.8
PATHOFINDER	RealAccurate Quadruplex Corona-plus PCR Kit	1	1	100,00%	31.4	31.4	31.4
SEEGENE	Allplex 2019 n-CoV Assay	9	9	100,00%	29.6	28.9	31.4
TIB MOLBIOL	LightMix Modular SARS and Wuhan CoV N-gene	3	3	100,00%	37.3	35.0	37.5
		22	22	100,00%	29.6	26.2	37.5
<b>ORF1a</b>							
MIKROGEN	amplicube Coronavirus SARS-CoV-2 RUO	1	1	100,00%			
ROCHE DIAGNOSTICS	COBAS SARS-CoV-2 Test	6	6	100,00%	28.7	28.0	30.3
		7	7	100,00%	28.7	28.0	30.3
<b>ORF1ab</b>							
APPLIED BIOSYSTEMS	TaqMan 2019-nCoV Assay Kit v1	1	1	100,00%	29.6	29.6	29.6
		1	1	100,00%	29.6	29.6	29.6

# Interim Evaluation

## Based on the Results of 112 out of 484 Laboratories

Reported by 14 April 2020 01:57 pm Berlin Time

**Sample: 340064 – SARS-CoV-2 positive (inactivated), 1 : 100 000 diluted**

Reagent	Testkit	Total (N=210)	positive (N=210)	Quota (100%)	Ct/Cp Median	Ct/Cp Min	Ct/Cp Max
<b>E-Gene</b>							
ALTONA DIAGNOSTIC	RealStar SARS-CoV-2 RT-PCR Kit 1.0	14	14	100,00%	29.1	26.5	32.5
OTHER MANUFACTURERS		2	2	100,00%			
IN HOUSE		7	7	100,00%	30.1	27.6	33.5
ELITech	GeneFinder COVID-19 Plus	2	2	100,00%	28.1	26.9	29.2
EURO IMMUN	EURORealTime SARS-CoV-2	1	1	100,00%	30.2	30.2	30.2
MIKROGEN	ampliCube Coronavirus SARS-CoV-2 RUO	1	1	100,00%			
Priv. Inst. f. Immunol. u. Mol.genetik	AmpliGnost CoV-2 E-Gen	3	3	100,00%	26.2	25.5	29.5
QIAGEN	QIAstat-Dx Respiratory 2019-nCoV Panel	2	2	100,00%	30.9	30.8	30.9
R-BIOPHARM	RIDA GENE SARS-CoV-2 RUO	10	10	100,00%	29.7	24.1	32.8
ROCHE DIAGNOSTICS	COBAS SARS-CoV-2 Test	13	13	100,00%	29.0	28.5	32.0
SEEGENE	Allplex 2019 n-CoV Assay	9	9	100,00%	27.9	26.7	30.7
TIB MOLBIOL	LightMix Modular SARS and Wuhan CoV E-gene	19	19	100,00%	29.4	27.3	31.5
		83	83	100,00%	29.05	24.1	33.5

# Interim Evaluation

## Based on the Results of 112 out of 484 Laboratories

Reported by 14 April 2020 01:57 pm Berlin Time

Sample: 340064 – SARS-CoV-2 positive (inactivated), 1 : 100 000 diluted

Reagent	Testkit	Total (N=210)	positive (N=210)	Quota (100%)	Ct/Cp Median	Ct/Cp Min	Ct/Cp Max
<b>RdRP-Gene</b>							
OTHER MANUFACTURERS		4	4	100,00%	33.5	28.5	38.5
IN HOUSE		6	6	100,00%	29.0	26.9	34.3
ELITECH	GeneFinder COVID-19 Plus	3	3	100,00%	30.4	30.4	30.8
SEEGENE	Allplex 2019 n-CoV Assay	13	13	100,00%	29.6	28.2	32.4
TIB MOLBIOL	LightMix Modular Wuhan CoV RdRP-gene	15	15	100,00%	33.5	29.5	41.4
		<b>41</b>	<b>41</b>	<b>100,00%</b>	<b>30.4</b>	<b>26.9</b>	<b>41.4</b>
<b>S-Gene</b>							
ALTONA DIAGNOSTIC	RealStar SARS-CoV-2 RT-PCR Kit 1.0	20	20	100,00%	28.3	25.0	31.5
OTHER MANUFACTURERS		1	1	100,00%			
APPLIED BIOSYSTEMS	TaqMan 2019-nCoV Assay Kit v1	1	1	100,00%	26.9	26.9	26.9
CERTEST BIOTEC	VIASURE SARS-CoV-2 S gene	5	5	100,00%	31.1	30.7	31.2
		<b>27</b>	<b>27</b>	<b>100,00%</b>	<b>28.6</b>	<b>25.0</b>	<b>31.5</b>
gene region not specified by participants							
ABBOTT	RealTime SARS-CoV-2	1	1	100,00%	16.0	16.0	16.0
ALTONA DIAGNOSTIC	RealStar SARS-CoV-2 RT-PCR Kit 1.0	1	1	100,00%	31.0	31.0	31.0
OTHER MANUFACTURERS		9	9	100,00%	30.8	27.8	34.7
BOSCH HEALTHCARE SOLUTIONS	Vivalytic VRI Multiplex Test	1	1	100,00%			
IN HOUSE		2	2	100,00%	25.5	25.5	25.5
FAST-TRACK DIAGNOSTICS	FTD SARS-CoV-2 assay	1	1	100,00%	28.7	28.7	28.7
GENEPROOF	SARS-CoV-2 PCR Kit	1	1	100,00%	25.4	25.4	25.4
GENESIG	Coronavirus COVID-19	3	3	100,00%	34.9	34.2	40.0
LUMINEX	NxTAG CoV	1	1	100,00%			
ROCHE DIAGNOSTICS	COBAS SARS-CoV-2 Test	5	5	100,00%	28.4	28.0	31.9
SEEGENE	Allplex 2019 n-CoV Assay	4	4	100,00%	30.2	28.0	31.1
		<b>29</b>	<b>29</b>	<b>100,00%</b>	<b>29.3</b>	<b>16.0</b>	<b>40.0</b>

# Interim Evaluation

## Based on the Results of 112 out of 484 Laboratories

Reported by 14 April 2020 01:57 pm Berlin Time

Sample No.	Sample properties	Expected qualitative result for SARS-CoV-2	Gene region	Correct results per reported results differentiated by gene region	Reported Ct/Cp-results differentiated by gene region median (min – max)
Sample 340059	<b>SARS-CoV-2 1 : 1 000 diluted<sup>\$</sup></b>	<b>positive</b>	E	83/83 (100%)	22.3 (17.7-29.4)
			N	22/22 (100%)	22.9 (17.9-30.9)
			ORF1a	7/7 (100%)	22.8 (21.7-28.7)
			ORF1ab	1/1 (100%)	20.9 (20.9-20.9)
			RdRP	41/41 (100%)	23.5 (20.5-34.4)
			S	27/27 (100%)	21.7 (18.0-24.8)
			n.s. <sup>§</sup>	29/29 (100%)	22.2 (10.4-33.0)
		<b>total</b>		<b>210/210 (100%)</b>	<b>22.6</b>
Sample 340064	<b>SARS-CoV-2 1 : 100 000 diluted<sup>\$</sup></b>	<b>positive</b>	E	83/83 (100%)	29.1 (24.1-33.5)
			N	22/22 (100%)	29.6 (26.2-37.5)
			ORF1a	7/7 (100%)	28.7 (28.0-30.3)
			ORF1ab	1/1 (100%)	29.6 (29.6-29.6)
			RdRP	41/41 (100%)	30.4 (26.9-41.4)
			S	27/27 (100%)	28.6 (25.0-31.5)
			n.s. <sup>§</sup>	29/29 (100%)	29.3 (16.0-40.0)
		<b>total</b>		<b>210/210 (100%)</b>	<b>29.4</b>
Sample 340060	<b>CoV OC43 1 : 2 500 diluted specificity control*</b>	<b>negative</b>	E	78/83 (94.0%)	-
			N	19/21 (90.5%) <sup>#</sup>	-
			ORF1a	7/7 (100%)	-
			ORF1ab	1/1 (100%)	-
			RdRP	36/39 (92.3%) <sup>#</sup>	-
			S	27/27 (100%)	-
			n.s. <sup>§</sup>	29/29 (100%)	-
		<b>total</b>		<b>197/210 (93.8%)<sup>#</sup></b>	-

# Thank you very much!

Univ.-Prof. i.R. Dr. Heinz Zeichhardt  
Charité-Universitätsmedizin Berlin

## Correspondence address

Prof. Dr. Heinz Zeichhardt

IQVD GmbH

Institut für Qualitätssicherung in der Virusdiagnostik

Potsdamer Chaussee 80

14129 Berlin, Germany

Tel.: +49-30-81054300

Fax: +49-30-81054303

Email: [Heinz.Zeichhardt@iqvd.de](mailto:Heinz.Zeichhardt@iqvd.de)

INSTAND EQA schemes in virus diagnostics

Web: [www.instandev.de](http://www.instandev.de)